



Specification

Bacteria-/Cell Count Standard MS 01 -2019

	BactoScan FC		
	MS A	MS B	Rel. MSB/MSA
Ref. value MS A	143 imp. ± 10%	283 imp. ± 10%	1,98

	Bacto nt		
	MS A	MS B	Rel. MSB/MSA
Ref. value MS A	174.000 IBC. ± 15%	345.000 IBC. ± 15%	1,98

Cell Count Standards (raw-milk) – No.01– Jan.2019	Cell Count milk "low"	Cell Count milk "high"	Cell Count milk "extra-high"
Reference value (<i>measuring range</i>)	174.000 cells/ml (156.000-192.000)	371.000 cells/ml (333.000 – 408.000)	522.000 cells/ml (469.000 –574.000)

Hergatz/Wangen 18.01.2016


 Dr. Josef Hüfner
