



Specification – reference values
Milk Standard - MS 07 – 2024

(A) BactoScan FC				
	MS A	MS B	MS C*)	RELATION MS B/MS A
Batch	MS 07	MS 07	MS 07	
Reference value MS 07 Impulses	158	468	1080	2,96
uncertainty of measurement -measuring range	142-174	421-515	972-1188	

(B) Bactocount (BC)				
	MS A	MS B	MS C*)	RELATION MS B/MS A
Batch	MS 07	MS 07	MS 07	
Reference value MS 07 Impulses	194	575	1328	2,96
uncertainty of measurement -measuring range	164-223	488-661	1196-1461	

Hergatz, den 04.07.2024



Dr. Josef Hüfner
Head of the institute

Durch die DAkkS nach DIN EN ISO 17034
akkreditierter Referenzmaterialhersteller



*) currently not part of the accreditation certificate

Erstellt von/am:	Hü / Mai 2014	Geändert von/am:	Bo / 16.06.23	Freigabe von/am:	Bo / 19.06.23
Geprüft von/am:	Hü / 11.09.19	Geprüft von /am	Hü / 19.06.23		
MIH		Specification BSC-FC,BC Milk Standard			Seite: 1 von 1 Version: 04