



Labor Dr.Hüfner GmbH

Specification – reference values
Cell Count Standard - MS 07 – 2024

Cow Milk – Cell Count Standards			
	Milk I “low”	Milk II “high”	Milk III “Extra-high”
Batch	07-2024	07-2024	07-2024
somatic cells x 1000	224	430	682
uncertainty of measurement -measuring range	202-246	387-473	614-750

Goat Milk – Special Cell Count Standards					
	Goat Milk “low”		Goat Milk “high”	Goat Milk “Extra-high”	
Batch	March 24	July 24	June 24	June 24	July 24
somatic cells x 1000	1050	1045	1.400	1.853	1.900
uncertainty of measurement -measuring range	893-1208	888-1201	1190-1610	1575-2131	1612-2181

Hergatz, 04.07.24

Dr. Josef Hüfner

Head of the Institute

Durch die DAkkS nach DIN EN ISO 17034
akkreditierter Referenzmaterialhersteller



Erstellt von/am:	Hü / Mai 2014	Geändert von/am:	Bo / 13.06.23	Freigabe von/am:	Bo / 19.06.23
Geprüft von/am:	Hü / 11.09.19	Geprüft von /am	Hü / 19.06.23		
		Specification Cell Count Standard			Seite: 1 von 1 Version: 04