

NEW – 12.06.2024

Specification – reference values

Cell Count Standard - MS 06 – 2024

Cow Milk – Cell Count Standards			
	Milk I “low”	Milk II “high”	Milk III “Extra-high”
Batch	06-2024	06-2024	06-2024
somatic cells x 1000	176	394	666
uncertainty of measurement -measuring range	158-194	354-433	599-733

Goat Milk – Special Cell Count Standards			
	Goat Milk “low”	Goat Milk “high”	Goat Milk “Extra-high”
Batch	March 24	June 24-NEW	June 24
somatic cells x 1000	1050	1.400	1.853
uncertainty of measurement -measuring range	893-1208	1190-1610	1575-2131

Hergatz, 06.06.24/12.06.24



Dr. Josef Hüfner

Head of the Institute

Durch die DAkkS nach DIN EN ISO 17034
akkreditierter Referenzmaterialhersteller



Erstellt von/am:	Hü / Mai 2014	Geändert von/am:	Bo / 13.06.23	Freigabe von/am:	Bo / 19.06.23
Geprüft von/am:	Hü / 11.09.19	Geprüft von /am	Hü / 19.06.23		
		Specification Cell Count Standard			Seite: 1 von 1 Version: 04